



ELSEVIER




IMAGE IN CARDIOLOGY

## Evaluation of left ventricular function using volumetric computed tomography in the clinical setting

Erasmo de la Peña-Almaguer<sup>a,\*</sup>, Alejandro R. Trevino<sup>b</sup>, Lilia Sierra-Galan<sup>c</sup>, José R. Azpiri-Lopez<sup>d</sup>, José L. Assad-Morell<sup>d</sup>, Jacobo Kirsch<sup>e</sup>, Su Min Chang<sup>f</sup>

<sup>a</sup> Cátedra de Cardiología, Departamento de Investigación Básica y Transferencia, Instituto de Cardiología y Medicina Vascular, Hospital Zambrano Helión, Tecnológico de Monterrey, Escuela de Medicina y Ciencias de la Salud, San Pedro Garza García, Nuevo León, Mexico

<sup>b</sup> Cátedra de Cardiología, Departamento de Cardiología, Hospital San José, Tecnológico de Monterrey, Escuela de Medicina y Ciencias de la Salud, Monterrey, Nuevo León, Mexico

<sup>c</sup> Imagen Cardiovascular, Departamento de Cardiología, Centro Médico ABC, México, D.F., Mexico

<sup>d</sup> Departamento de Cardiología, Hospital Christus Muguerza Alta Especialidad, Universidad de Monterrey, Escuela de Medicina, Monterrey, Nuevo León, Mexico

<sup>e</sup> Cardiovascular Imaging, Department of Radiology, Cleveland Clinic Florida Hospital, Cleveland Clinic, Weston, FL, United States

<sup>f</sup> Cardiovascular Imaging, Department of Cardiology, DeBakey Heart and Vascular Center, The Methodist Hospital, Weill Cornell College of Medicine, Houston, TX, United States

Received 29 April 2015; accepted 20 May 2015

### KEYWORDS

Computed tomography;  
Cardiovascular magnetic resonance;  
Ejection fraction;  
Mexico

### Abstract

**Objective:** To correlate the left ventricular parameters obtained with 64-slice Volumetric Computed Tomography (VCT) with those obtained with the reference standard, cardiovascular magnetic resonance (CMR) imaging.

**Methods:** VCT and a 3.0 T MRI scanner were used. Results from both studies were independently evaluated by two cardiologists. A linear correlation and a paired Student's *t* test were used to analyze the data with a *P*<0.05 being considered significant.

**Results:** Thirty consecutive patients were evaluated with VCT and CMR. The left ventricular indices for CMR and VCT were, respectively, mass  $86.4 \pm 25.8$  vs.  $82.7 \pm 27.6$  g (*P*=0.31); ESV  $45.5 \pm 27.8$  vs.  $48.7 \pm 40.4$  ml (*P*=.405); EDV  $101.3 \pm 32.7$  vs.  $105.1 \pm 44.0$  ml (*P*=0.475); SV  $55.9 \pm 16.1$  vs.  $56.8 \pm 15.6$  ml (*P*=0.713); LVEF  $57.5 \pm 13.2\%$  vs.  $56.9 \pm 12.4\%$  (*P*=0.630). No differences in intraobserver variability for both methods were found, CT  $r=0.96$ ,  $r^2=0.92$  *P*<0.0001 and MR  $r=0.96$   $r^2=0.93$  *P*<0.0001. There was no significant statistical difference in the presence of artifacts.

\* Corresponding author at: Centro Médico Zambrano-Helión, Batallón de San Patricio 112, Edificio Instituto de Cardiología y Medicina Vascular 2do piso, Real de San Agustín, San Pedro Garza García, Nuevo León 66278, Mexico. Tel.: +52 81 88880500; fax: +52 81 88880500.

E-mail address: [edelapena@itesm.mx](mailto:edelapena@itesm.mx) (E. de la Peña-Almaguer).

**Conclusion:** There is a close correlation between CMRI and VCT in the evaluation of LV function. VCT is as useful as 3 T CMR, and could be incorporated as another resource for evaluating LV function.

© 2015 Instituto Nacional de Cardiología Ignacio Chávez. Published by Masson Doyma México S.A. All rights reserved.

## PALABRAS CLAVE

Tomografía  
computarizada;  
Resonancia  
magnética cardiaca;  
Fracción de eyección;  
México

## Evaluación de la función ventricular con tomografía computada volumétrica en el ámbito clínico

### Resumen

**Objetivo:** Correlacionar los parámetros del ventrículo izquierdo obtenidos con la tomografía computarizada volumétrica (TCV) de 64 cortes con el estándar de referencia, la resonancia magnética cardiaca (RMC).

**Métodos:** Se usó una TCV de 64 cortes y una RMC de 3.0 Teslas. Ambos estudios fueron interpretados de manera independiente por dos cardiólogos. Los datos se analizaron con la prueba t de Student y una correlación lineal, resultando significativo valores de  $P < 0.05$ .

**Resultados:** Treinta pacientes consecutivos fueron evaluados con TCV y RMC. Los índices del ventrículo izquierdo mediante RMC y TCV fueron, respectivamente, masa  $86.4 \pm 25.8$  vs.  $82.7 \pm 27.6$  g ( $P = 0.31$ ); volumen telesistólico  $45.5 \pm 27.8$  vs.  $48.7 \pm 40.4$  ml ( $P = 0.405$ ); volumen telediastólico  $101.3 \pm 32.7$  vs.  $105.1 \pm 44.0$  ml ( $P = 0.475$ ); volumen latido  $55.9 \pm 16.1$  vs.  $56.8 \pm 15.6$  ml ( $P = 0.713$ ); fracción de eyección de ventrículo izquierdo  $57.5 \pm 13.2\%$  vs.  $56.9 \pm 12.4\%$  ( $P = 0.630$ ). No hubo diferencias entre la variabilidad intraobservador para ambos métodos, CT  $r = 0.96$   $r^2 = 0.92$   $P < 0.0001$  y MR  $r = 0.96$   $r^2 = 0.93$ ,  $P < 0.0001$ . No se encontraron diferencias significativas en la presencia de artefactos.

**Conclusión:** Hay correlación estrecha entre RMC y TCV en la evaluación de la función ventricular. La TCV es tan útil como la RMC y pudiera incorporarse como otro recurso para valorar la función ventricular.

© 2015 Instituto Nacional de Cardiología Ignacio Chávez. Publicado por Masson Doyma México S.A. Todos los derechos reservados.

## Introduction

The evaluation of left ventricular function and mass has both prognostic and therapeutic implications; therefore, pharmacological, interventional or surgical interventions of left ventricular dysfunction require an adequate tool for staging and follow-up.<sup>1,2</sup>

Cardiovascular magnetic resonance (CMR) at 1.5 T has been proven reliable for the evaluation of left ventricular (LV) function and is still used in many centers for cardiac assessment. It is currently accepted as the standard of reference for the evaluation of LV volumes, ejection fraction, and mass.<sup>2,3</sup> The use of CMR at 3 T has increased over the last few years. 3-T CMR provides a higher magnetic field, which translates into an increased contrast to noise ratio and signal to noise ratio, resulting in a clearer image of anatomical structures. It also allows faster image acquisition and a shorter repetition time. However, technical difficulties, such as compromise in B0 and B1 homogeneity, increased artifact resulting in signal loss, and increased specific absorption rate, which may approach the upper limit, have arisen from this technique.<sup>4-6</sup>

Multi-Slice Cardiovascular Computed Tomography using 16 slices or higher is another tool that has proved to be

effective in the evaluation of coronary artery disease. This method allows concurrent evaluation of coronary anatomy and ventricular function, showing accurate values when compared to the reference standard, CMR.<sup>7-10</sup>

The objectives of this study are to evaluate the correlation between LV parameters, global and regional wall motion abnormalities, and the presence of artifacts using 3-T CMR and 64-slice Volumetric Computed Tomography (VCT), in patients with known or suspected coronary artery disease.

## Materials and methods

### Subjects

Thirty patients (23 men, aged  $54.3 \pm 8$  years) who presented with known or suspected coronary artery disease between April 2006 and July 2007 and who were referred for computed tomography coronary angiography and CMR underwent examination on a 3 T MR scanner and cardiac VCT. Patient preparation and screening for contraindications were essential to ensure patient safety and optimal quality studies. Exclusion criteria for CT examination were renal insufficiency (serum creatinine  $> 1.5$  mgs/dl), atrial fibrillation or another irregular rhythm. Patients with an unstable

condition or an intervention between the different examinations were excluded. All patients provided written informed consent for contrast administration and the hospital ethics and research committee approved the study protocol.

## Magnetic resonance imaging

### Image acquisition

A 3-T MRI system using a cardiac 8 channel coil, with 23 to 50 mT/m, slew rate of 80–150 T/m/s was used (Signa Excite HD, GE Medical Systems, Milwaukee, WI). Cine-MRI was performed in short axis (SAX) (6–8 slices, steady-state free precession sequences) with additional horizontal long axis (HLA) and vertical long axis (VLA) slices, using a 10-mm thickness, no spacing, Flip Angle 40, Field of View (FOV) 38–44 cm, Partial FOV of 0.75, phase 320, frequency 224, 1 NEX, 24 vps. Offline analysis of Global functional parameters were obtained with commercially available software (MASS analysis v6.1, Medis, NL) and as needed, corrected by using manual definition of the LV borders, applying the same method for the CT studies. Also, measurements were obtained twice for both methods to analyze interobserver variability of LV parameters.

### Myocardial delayed imaging

Myocardial Delayed Enhanced (MDE) images were obtained in short axis (6–8 slices), HLA, and VLA 10 min after 0.2 mmol kg<sup>-1</sup> Gd-DTPA at a 3 ml s<sup>-1</sup> rate was administered. An inversion recovery-fast-gradient-echo (IR-FGRE) sequence was used with a slice thickness of 10 mm, no spacing, flip angle 20°, whole gradient, Partial FOV 44 cm, NEX 2, Phase 128, Frequency 128, 24 vps, with an inversion time of 175–250 ms, adjusted to obtain proper myocardial saturation. Myocardial necrosis was defined as the region of the myocardial wall in which the persistence of the contrast agent was observed in two contiguous slices in the area of at least one coronary territory. A transmural score was obtained (thickness of the contrast enhanced wall divided by the thickness of the whole myocardial wall × 100%) and classified according to its extension from such images and classified as follows: 1–25%, 26–50%, 51–75% and >75%, for the assessment of abnormal regions and to determine whether a segmental wall motion abnormality (WMA) was due to an artifact, an infarct or both.

### Cardiac tomography acquisition

A 64-slice VCT (GE Medical Systems, Milwaukee, WI) was used for ECG-gated cardiac data acquisition. After a right antecubital 18 Fr vein access was obtained, intravenous beta blocker was administered in all patients with no oral beta blocker or contraindication (metoprolol 5–30 mg IV) to reduce heart rate to <65 beats/min. Sequential images were acquired at the level of the aorta and pulmonary artery using 20 ml of a non-ionic contrast (Iopamidol, 370, Bracco Diagnostics, Inc, Monroe Township, NJ) at 5–7 ml s<sup>-1</sup>, followed by a saline bolus of 20 ml for timing of the arrival

of contrast; 3 s were added to the time delay. The scanner parameters were cardiac segment or cardiac burst, retrospective ECG-gating with ECG modulation for dose reduction, gantry rotation of 350 ms, gantry tilt 0, slice thickness of 0.625 mm × 0.625 mm, speed 6.0, pitch 0.3:1, SFOV Large, Kv 100–120, ECG Modulated mA of 400–740.

Images were acquired during apnea after hyperventilation. A retrospective reconstruction of the cardiac cycle was performed using 10% space intervals to obtain a complete cardiac cycle. Offline analysis of the images was performed at a workstation (Advantage Windows v4.2, GE Medical Systems, Milwaukee, WI). These were 6–10 SAX, HLA, and VLA obtained in a manner similar to that of cardiac MRI, with a 10-mm slice thickness and a 0.5-mm interval.

### Left ventricular function parameters

#### Global parameters

Global left ventricular ejection fraction (EF), end-diastolic volume (EDV), end-systolic volume (ESV), stroke volume (SV) and myocardial mass parameters were obtained and compared to those of VCT.

#### Assessment of wall motion abnormality

For VCT and CMR, cine loops of short axis slices (basal, mid-ventricular, apical,) as well as the 2 and 4 chamber orientation and graded using a 17-segment LV model according to the American Heart Association/American College of Cardiology guidelines were evaluated by 2 independent, blinded, experienced cardiologists, who had no knowledge of clinical information.<sup>11</sup> Images were scored as follows: normal = 0, hypokinesis = 1, akinesis = 2 and dyskinesis = 3. The WMA was also matched to its corresponding segment on the MDE images.

#### Evaluation of Image quality and Artifacts MRI

MRI and CT artifacts were graded as the amount of LV myocardial wall not visualized in its entirety due to artifacts in each of the 17 segments as follows: 0 = normal, 1 = myocardium not visualized < 25%, 2 = myocardium not visualized > 26–50%, 3 = myocardium not visualized > 51–75%, 4 = myocardium not visualized > 76–100%.

#### Statistical analysis

Data are expressed as mean values ± standard deviation (SD). Comparisons between groups were assessed by a Two-sample *t* test and chi-square analysis for continuous and categorical variables. The correlation between continuous variables was assessed by Pearson's test, and Bland–Altman analysis was used to compare the quantitative data of CT with the reference standard CMR. The degree of agreement was expressed as the mean difference, limits of agreement (2 SD), and standard error of the mean differences. Kappa analyses of the global agreement between two observers in image quality and regional wall motion scores between VCT and CMR were performed and valued as follows:

**Table 1** Patient demographics.

Variable	n (%)
Total	30
Age, years	55 ± 12
Male	22 (73)
Mean time between scans	1.2 days
Comorbidities:	
Ischemic heart disease	21 (70)
Dilated cardiomyopathy	4 (13.3)
Hypertension	3 (10)
Hyperthrophic cardiomyopathy	1 (3.3)
Pericarditis	1 (3.3)
MDE + MR	12

MDE + MR, myocardial delayed enhanced + magnetic resonance

0–0.2 low, 0.21–0.40 moderate, 0.41–0.60 substantial, 0.61–0.80 good, and ≥0.81 perfect agreement. The correlations between HR and the percent difference of EF, EDV, ESV (percent difference (%)) = [(CT – CMR)/CMR] – 100%, and summed scores between VCT and CMR, were additionally assessed by linear regression and correlation coefficient analyses. Data are expressed as means ± standard deviation (SD). A *P* value <0.05 was considered statistically significant.

## Results

Thirty patients with a mean age of 55 ± 12 years were included; 22 were men (73%) and 8 were women (27%). The mean time between scans was 1.2 days. No intervention was performed nor major hemodynamically or significant clinical events occurred during such scans. Clinical indications for the scans were ischemic heart disease, 21 (70%), dilated cardiomyopathy, 4 (13.3%), hypertension, 3 (10%), hyperthrophic cardiomyopathy, 1 (3.3%), and pericarditis, 1 (3.3%). (Table 1)

### Left ventricular function

Overall, a strong correlation between LV function values for both methods was obtained. An overestimation of mass for CT was noted, although this was not statistically significant. The results for left ventricular indices were (for MRI and CT, respectively) mass 86.4 ± 25.8 vs CT 82.7 ± 27.6 g (*P* = 0.31), ESV 45.5 ± 27.8 vs. 48.7 ± 40.4 ml (*P* = 0.405), EDV 101.3 ± 32.7 vs 105.1 ± 44.0 ml (*P* = 0.475), SV 55.9 ± 16.1 vs 56.8 ± 15.6 ml (*P* = 0.713); LVEF 57.5 ± 13.2% vs 56.9 ± 12.4% (*P* = 0.630). (Table 2) No significant differences were found in intraobserver variability for either method CT *r* = 0.96 *r*<sup>2</sup> = 0.92, *P* < 0.0001 and MR *r* = 0.96 *r*<sup>2</sup> = 0.93, *P* < 0.0001 (Fig. 1).

The rank correlation analysis included 1020 segments assessed by two experienced observers. Observers analyzed segmental wall motion abnormalities as previously described in the methods section. A strong correlation for both methods was found among observers, Spearman  $\rho$  = 0.554; *P* < 0.001 Kendall  $\tau_c$  = 0.211; *P* < 0.001. (Table 3)

Interobserver variability assessing wall motion was low for both methods with a slight difference in documenting

**Table 2** Global left ventricular function assessed by cardiovascular magnetic resonance (CMR) and computed tomography.

Parameter	CT (n = 30)	CMR (n = 30)	<i>P</i>
Mass, g	86.4 ± 25.8	82.7 ± 27.6	0.311
ESV, ml	45.5 ± 27.8	48.7 ± 40.4	0.405
EDV, ml	101.3 ± 32.7	105.1 ± 44.0	0.475
SV, ml	55.9 ± 16.1	56.8 ± 15.6	0.713
LVEF, %	57.5 ± 13.2	56.9 ± 12.4	0.63

CMR, cardiovascular magnetic resonance; VCT, volumetric computed tomography; EDV, end diastolic volume; ESV, end systolic volume; LVEF, left ventricular ejection fraction; SV, stroke volume.


normal versus hypokinesis; CT had an observed concordance = 0.796, an expected concordance = 0.701  $\kappa$  = 0.3178; 95% CI = 0.2302–0.4350; *P* < 0.00001. (Table 4) When compared to VCT, CMR had fewer non-interpretable studies, but had a greater number of variability between observers; observed concordance = 0.78, expected concordance = 0.66,  $\kappa$  = 0.3496; 95% CI = 0.2702–0.4290 (Table 5). The incidence of artifacts was not statistically different among methods, where most artifacts involved less than 25–50% of subendocardial border visualization.

## Discussion

The use of high field magnetic resonance imposes certain benefits over the conventional standard of reference, CMR at 1.5 T; however, initial results have shown a higher incidence of artifacts with 3 T. High-field CMR offers a better temporal and spatial resolution, which translates into a faster anatomical evaluation due to faster sequences. The resonance spins are increased, providing robust fat saturation. There is a better response on blood oxygen level dependency (BOLD) effect for fMRI, as well as a greater peak dispersion, making it ideal for spectroscopy studies.<sup>12</sup> However, a little or no data are available on 3 T for the evaluation of LV function and WMA, as compared to data analyzed by CT and other modalities.<sup>8–10</sup>

Clinical scenarios where CMR is used can be diverse, including research in normal volunteers, in which no pathology is present and normal hemodynamics and breathing patterns are observed. Imaging patients with pathology may pose some image acquisition problems when compared to normal subjects. Therefore, the limitations of this study include inherent variables due to pathologies, the presence of increased intravascular volumes due to contrast injection during the acquisition of CT datasets, where a non-significant increase in LV volumes was noted in those of CMR, and the use of beta blockers, which may affect stroke volume and exert a negative inotropic effect.

We found a close correlation in the evaluation of global LV function systolic indices; however, in our study 64-slice VCT showed a non-significant tendency to underestimate LV mass. Regional wall motion assessment had a highly significant correlation between both methods, although 75% of the segments analyzed were normal. Interestingly, wall motion of septal segments does not correlate as well as that of the rest of the LV segments. Interobserver variability for wall



**Figure 1** A strong correlation between computed tomography (CT) and cardiovascular magnetic resonance (CMR) was seen when assessing left ventricular mass, end systolic volume (ESV), end diastolic volume (EDV) and left ventricle ejection fraction (LVEF).

**Table 3** Correlation between volumetric computed tomography and cardiovascular magnetic resonance when evaluating wall motion abnormalities. Spearman  $\rho = 0.554$ ;  $p < 0.001$  Kendall  $\tau - c = 0.211$ ;  $p < 0.001$ .

CMR-17 segments WMA						
	Non interpretable	Normal	Hypokinesis	Akinesis	Dyskinesis	
<i>VCT-64-17 segments WMA</i>						
Non interpretable	0	4	1	2	0	7
Normal	2	764	65	15	0	846
Hypokinesis	0	44	39	10	2	95
Akinesis	0	10	12	33	9	64
Dyskinesis	0	0	1	1	6	8
Total	2	822	117	61	17	1020

CMR, cardiovascular magnetic resonance; VCT, volumetric computed tomography; WMA, wall motion abnormalities.

**Table 4** Interobserver variability of wall motion abnormalities with volumetric computed tomography.

WMA-VCT-64 Observer 1						
	Non interpretable	Normal	Hypokinesis	Akinesis	Dyskinesis	
<i>WMA-VCT-64 Observer 2</i>						
Non interpretable	0	0	0	0	0	0
Normal	3	382	27	8	0	420
Hypokinesis	0	43	6	8	0	59
Akinesis	1	2	2	18	0	23
Dyskinesis	1	0	0	7	0	8
Total	7	427	35	41	0	510

CT, volumetric computed tomography; WMA, wall motion abnormalities.

**Table 5** Interobserver variability of wall motion abnormalities with cardiovascular magnetic resonance.

	WMA CMR Observer 1					
	Non interpretable	Normal	Hypokinesis	Akinesis	Dyskinesis	Total
<b>WMA CMR Observer 2</b>						
Non interpretable	0	0	0	0	0	0
Normal	0	372	23	2	0	397
Hypokinesis	2	44	16	14	0	76
Akinesis	0	8	1	11	0	20
Dyskinesis	0	2	1	14	0	17
Total	2	426	41	41	0	510

CMR, cardiovascular magnetic resonance; WMA, wall motion abnormalities.

motion assessed by 64-VCT goes beyond that expected by chance. No significant difference was found in the presence of artifacts when assessing global and regional LV systolic function by 64-VCT.

## Conclusion

In this study, 3-T CMR was comparable to 64-VCT in the assessment of LVEF, WMA, volumes, and LV mass without a significant incidence of artifacts. This makes 3-T CMR an excellent option to evaluate these parameters.

## Responsabilidades éticas

**Protección de personas y animales.** Los autores declaran que para esta investigación no se han realizado experimentos en seres humanos ni en animales.

**Confidencialidad de los datos.** Los autores declaran que en este artículo no aparecen datos de pacientes.

**Derecho a la privacidad y consentimiento informado.** Los autores han obtenido el consentimiento informado de los pacientes y/o sujetos referidos en el artículo. Este documento obra en poder del autor de correspondencia.

## Funding

This work was partially supported by Endowed Chair in Cardiology-Tec de Monterrey 0020CAT131.

## Conflict of interest

The authors declare no conflict of interest.

## References

1. Davis KB, Chaitman B, Ryan T, Bittner V, Kennedy JW. Comparison of 15-year survival for men and women after initial medical or surgical treatment for coronary artery disease: a CASS registry study. *Coronary Artery Surgery Study. J Am Coll Cardiol.* 1995;25:1000–9.
2. Epstein FH. MRI of left ventricular function. *J Nucl Cardiol.* 2007;14:729–44.
3. Fieno DS, Thomson LE, Slomka PJ, et al. Rapid assessment of left ventricular segmental wall motion, ejection fraction, and volumes with single breath-hold, multi-slice TrueFISP MR imaging. *J Cardiovasc Magn Reson.* 2006;8:435–44.
4. England B, Lee A, Tran T, et al. Magnetic resonance criteria for future trials of cardiac resynchronization therapy. *J Cardiovasc Magn Reson.* 2005;7:827–34.
5. Yang PC, Nguyen P, Shimakawa A, et al. Spiral magnetic resonance coronary angiography – direct comparison of 1.5 Tesla vs. 3 Tesla. *J Cardiovasc Magn Reson.* 2004;6:877–84.
6. Oshinski JN, Delfino JG, Sharma P, Gharib AM, Pettigrew RI. Cardiovascular magnetic resonance at 3.0 T: current state of the art. *J Cardiovasc Magn Reson.* 2010;12:55.
7. Abbara S, Chow BJ, Pena AJ, et al. Assessment of left ventricular function with 16- and 64-slice multi-detector computed tomography. *Eur J Radiol.* 2007.
8. De la Peña-Almaguer E, Azpíri Lopez JR, González-Camid Fde J, et al. Evaluation of left ventricular function with a 16-slice multidetector tomograph (MDCT-16): correlation with cardiovascular magnetic resonance imaging]. *Arch Cardiol Mex.* 2005;75:55–60.
9. Schepis T, Gaemperli O, Koepfli P, et al. Comparison of 64-slice CT with gated SPECT for evaluation of left ventricular function. *J Nucl Med.* 2006;47:1288–94.
10. Wu YW, Tadamura E, Kanao S, et al. Left ventricular functional analysis using 64-slice multidetector row computed tomography: comparison with left ventriculography and cardiovascular magnetic resonance. *Cardiology.* 2007;109:135–42.
11. Cerqueira MD, Weissman NJ, Dilsizian V, et al. Standardized myocardial segmentation and nomenclature for tomographic imaging of the heart. A statement for healthcare professionals from the Cardiac Imaging Committee of the Council on Clinical Cardiology of the American Heart Association. *Int J Cardiovasc Imaging.* 2002;18:539–42.
12. Jahnke C, Gebker R, Manka R, et al., Schnackenburg B, Fleck E, Paetsch I. Navigator-gated 3D blood oxygen level-dependent CMR at 3.0-T for detection of stress-induced myocardial ischemic reactions. *JACC Cardiovasc Imaging.* 2010;3: 375–84.